


TCVGH Temperature Log



2022/6/29 下午 07:00:00 2022/7/4 下午 07:00:00 2022/7/9 下午 07:00:00 2022/7/14 下午 07:00:00 2022/7/19 下午 07:00:00 2022/7/24 下午 07:00:00 2022/7/29 下午 07:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C011
Serial Number	1803-0048
Product Code	EI-IN-D-32-L
Start	2022/6/29 下午 07:00:00
Finish	2022/8/1 上午 10:30:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	3.8 °C
Highest	4.7 °C
Lowest	3.3 °C

Yu Chen Su
01 Aug 2022