

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C001 (Model: TIT / TR200)

Thermograph: ESCORT,1043-0065

Date (dd-mmm-yy)	Current (AM)			Current (PM)			Max. & Min.			Comments
	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
01-Feb-17	08:27	4.6	NTH	13:00	4.4	NTH	5.2	3.4	NTH	
02-Feb-17	08:07	4.1	NTH	19:16	4.4	NTH	5.2	3.4	NTH	
03-Feb-17	08:20	4.6	NTH	17:00	5.1	STW	6.4	3.7	NTH	
06-Feb-17	10:00	5.3	NTH	17:15	5.3	STW	6.2	4.2	NTH	
07-Feb-17	08:16	4.7	NTH	17:05	4.7	STW	5.8	4.2	NTH	
08-Feb-17	08:20	4.7	NTH	17:10	5.2	STW	5.9	4.3	NTH	
09-Feb-17	08:13	4.4	NTH	17:15	5.0	STW	6.2	4.1	NTH	
10-Feb-17	08:30	4.7	NTH	17:05	5.1	STW	6.3	4.4	NTH	
13-Feb-17	08:12	4.6	NTH	17:00	4.9	STW	6.5	4.3	NTH	
14-Feb-17	08:03	5.9	NTH	17:25	5.2	NTH	6.1	4.3	NTH	
15-Feb-17	08:15	5.0	NTH	18:00	5.3	NTH	5.9	4.3	NTH	
16-Feb-17	08:45	5.1	NTH	17:29	4.9	NTH	6.3	4.3	NTH	
17-Feb-17	08:00	4.5	NTH	17:05	4.5	NTH	6.2	4.4	NTH	
20-Feb-17	08:57	5.4	NTH	18:03	4.4	NTH	6.1	4.4	NTH	
21-Feb-17	08:06	4.7	NTH	17:05	4.9	STW	6.1	4.4	NTH	
22-Feb-17	08:17	4.9	NTH	17:23	4.7	NTH	6.3	4.1	NTH	
23-Feb-17	08:50	4.9	NTH	17:00	4.6	STW	5.9	4.1	NTH	
24-Feb-17	08:48	4.4	NTH	17:00	4.2	STW	5.9	4.2	NTH	
27-Feb-17	08:27	4.4	NTH	17:11	5.1	NTH	6.0	4.2	NTH	

NTH 07 Mar 2017