

# Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C001 ( Model: TIT / TR200 )

Thermograph: ESCORT,1043-0065

Date (dd-mmm-yy)	Current (AM)			Current (PM)			Max. & Min.			Comments
	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
01-Mar-17	08:10	4.5	NTH	17:15	4.4	STW	6.0	4.3	NTH	
02-Mar-17	09:23	4.6	NTH	17:22	5.2	NTH	5.9	4.2	NTH	
03-Mar-17	08:26	5.1	NTH	16:58	4.3	STW	6.2	4.2	NTH	
06-Mar-17	08:25	4.4	NTH	17:00	4.3	STW	5.9	4.2	NTH	
07-Mar-17	08:12	4.3	NTH	17:10	4.2	STW	5.7	4.2	NTH	
08-Mar-17	09:30	4.6	NTH	17:00	4.9	NTH	5.4	4.0	NTH	
09-Mar-17	09:20	5.3	NTH	17:40	4.2	STW	6.2	4.1	NTH	
10-Mar-17	08:13	5.1	NTH	17:30	4.3	STW	6.2	4.2	NTH	
13-Mar-17	08:05	5.1	STW	17:30	5.3	STW	5.8	4.2	NTH	
14-Mar-17	08:05	4.8	STW	17:25	4.8	STW	5.8	4.2	NTH	
15-Mar-17	08:05	5.4	STW	17:40	4.4	STW	5.4	4.2	NTH	
16-Mar-17	08:05	5.0	STW	17:25	4.7	STW	6.1	4.3	NTH	
17-Mar-17	08:05	4.4	STW	17:20	4.4	STW	5.9	4.2	NTH	
20-Mar-17	08:40	4.8	NTH	17:05	4.7	STW	6.2	4.2	NTH	
21-Mar-17	09:25	4.6	NTH	17:45	4.6	NTH	6.1	4.2	NTH	
22-Mar-17	08:10	4.6	NTH	18:15	4.5	STW	6.3	4.2	NTH	
23-Mar-17	08:10	4.4	NTH	16:55	4.9	STW	6.0	4.3	NTH	
24-Mar-17	10:05	4.6	NTH	17:05	4.9	STW	6.2	4.2	NTH	
27-Mar-17	08:40	5.2	NTH	17:05	4.8	STW	6.0	4.1	NTH	
28-Mar-17	09:30	4.7	NTH	16:55	4.4	STW	6.2	4.1	NTH	
29-Mar-17	09:25	4.9	NTH	17:30	4.3	STW	6.1	4.2	NTH	
30-Mar-17	08:10	5.0	NTH	17:05	4.4	STW	5.8	4.0	NTH	
31-Mar-17	10:05	4.3	NTH	17:15	4.6	NTH	6.7	4.0	NTH	

Nair NTH 05 Apr 2017