

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C001 (Model: TIT / TR200)

Thermograph: ESCORT,1043-0065

Date (dd-mmm-yy)	Current (AM)			Current (PM)			Max. & Min.			Comments
	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
01-May-17	09:00	5.4	NTN	17:00	4.3	STHW	7.6	4.2	NTN	
02-May-17	09:18	5.5	NTN	18:05	5.4	NTN	5.6	4.2	NTN	
03-May-17	08:38	4.7	NTN	17:00	5.6	STHW	5.7	4.2	NTN	
04-May-17	09:15	5.1	NTN	16:55	4.5	STHW	6.2	4.1	NTN	
05-May-17	10:00	4.4	NTN	17:30	4.5	NTN	6.5	4.2	NTN	
08-May-17	08:40	5.4	NTN	17:00	4.4	STHW	5.9	4.3	NTN	
09-May-17	09:10	5.6	NTN	17:00	5.7	STHW	6.6	4.3	NTN	
10-May-17	09:41	4.5	NTN	17:25	5.2	STHW	6.0	4.4	NTN	
11-May-17	09:08	4.4	NTN	17:00	4.8	STHW	6.7	4.4	NTN	
12-May-17	08:15	4.4	NTN	17:00	4.4	STHW	6.0	3.8	NTN	
15-May-17	09:05	4.2	NTN	18:00	4.4	NTN	5.6	3.8	NTN	
16-May-17	09:35	5.8	NTN	17:40	4.7	NTN	6.3	4.3	NTN	
17-May-17	09:37	5.2	NTN	17:57	5.1	NTN	6.0	4.3	NTN	ESCORT(1043-0065) was sent for calibration on 19 May 2017. ESCORT (118-0020) as back up.
18-May-17	09:05	4.5	NTN	17:20	4.3	STHW	6.1	4.2	NTN	
19-May-17	08:30	4.6	NTN	17:20	5.1	NTN	5.9	4.0	NTN	
22-May-17	08:15	4.7	NTN	17:40	4.7	NTN	5.7	4.1	NTN	NTN 19 May 2017
23-May-17	09:10	5.3	NTN	17:40	5.3	NTN	5.6	4.3	NTN	
24-May-17	09:05	5.3	NTN	17:20	4.2	NTN	6.1	4.0	NTN	
25-May-17	09:19	4.5	NTN	18:10	5.0	NTN	5.5	4.0	NTN	
26-May-17	08:24	4.3	NTN	17:00	4.9	STHW	5.9	4.2	NTN	
29-May-17	08:30	4.4	NTN	18:04	4.7	NTN	6.3	3.9	NTN	
31-May-17	09:05	4.2	NTN	17:00	4.7	STHW	6.3	3.9	NTN	

Naiting 01 Jun 2017