

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C001 (Model: TIT / TR200)

Thermograph: ESCORT,1043-0065

Date (dd-mmm-yy)	Current (AM)			Current (PM)			Max. & Min.			Comments
	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
01-Sep-17	09:10	5.1	~T~	16:50	4.2	STW	6.2	3.6	~T~	
04-Sep-17	08:05	4.1	~T~	17:05	3.8	STW	6.4	3.7	~T~	
05-Sep-17	09:35	3.7	~T~	16:50	3.9	STW	6.3	3.6	~T~	
06-Sep-17	08:16	4.9	~T~	17:46	4.0	~T~	5.2	3.7	~T~	
07-Sep-17	08:15	3.8	~T~	17:05	4.8	STW	5.9	3.7	~T~	
08-Sep-17	08:15	3.7	~T~	17:30	3.7	STW	5.2	3.6	~T~	
11-Sep-17	08:35	3.8	~T~	17:58	4.2	~T~	5.7	3.7	~T~	
12-Sep-17	09:20	3.9	~T~	18:45	3.9	~T~	6.2	3.9	~T~	
13-Sep-17	08:58	4.0	~T~	17:00	4.8	STW	6.2	3.7	~T~	
14-Sep-17	09:00	3.8	~T~	17:00	3.9	STW	5.3	3.7	~T~	
15-Sep-17	08:08	4.5	~T~	17:07	4.2	STW	5.3	3.7	~T~	
18-Sep-17	08:30	4.5	~T~	17:00	5.1	STW	5.2	3.6	~T~	
19-Sep-17	09:16	4.4	~T~	17:24	5.0	STW	6.0	3.6	~T~	
20-Sep-17	08:52	5.2	~T~	17:05	3.9	STW	5.3	3.6	~T~	
21-Sep-17	09:19	3.7	~T~	17:25	3.8	STW	5.2	3.6	~T~	
22-Sep-17	09:35	4.9	~T~	17:00	3.9	STW	6.5	3.5	~T~	
25-Sep-17	08:26	5.1	~T~	17:00	3.3	STW	5.2	2.6	~T~	
26-Sep-17	09:30	4.8	~T~	17:05	4.9	STW	5.1	3.3	~T~	
27-Sep-17	09:05	4.8	~T~	17:40	4.4	~T~	6.6	3.2	~T~	
28-Sep-17	09:15	3.3	~T~	17:00	3.4	~T~	6.0	3.1	~T~	
29-Sep-17	09:23	4.3	~T~	17:05	5.2	~T~	6.0	3.5	~T~	

Na - L My 02 Oct 2017