


# TCVGH Temperature Log



Sensor Name	Internal
Legend	
Description	PHCL-C003
Serial Number	1803-0048
Product Code	EI-IN-D-32-L
Start	2018/12/28 下午 12:00:00
Finish	2019/2/1 下午 03:20:00
Specification	2.0 to 8.0 °C
Average	3.9 °C
Highest	4.9 °C
Lowest	2.8 °C

Na-zhy 01 Feb 2019