


TCVGH Temperature log



2019/2/27 上午 10:00:00 2019/3/4 上午 10:00:00 2019/3/9 上午 10:00:00 2019/3/14 上午 10:00:00 2019/3/19 上午 10:00:00 2019/3/24 上午 10:00:00 2019/3/29 上午 10:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C003
Serial Number	1803-0048
Product Code	EI-IN-D-32-L
Start	2019/2/27 上午 10:00:00
Finish	2019/4/1 上午 08:50:00
Specification	2.0 to 8.0 °C
Average	4.2 °C
Highest	5.3 °C
Lowest	3.4 °C

Yunwenwang
01 Apr 2019