


# TCVGH Temperature Log



2019/5/31 下午 01:00:00    2019/6/5 下午 01:00:00    2019/6/10 下午 01:00:00    2019/6/15 下午 01:00:00    2019/6/20 下午 01:00:00    2019/6/25 下午 01:00:00    2019/6/30 下午 01:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C003
Serial Number	1715-0143
Product Code	EI-IN-D-32-L
Start	2019/5/31 下午 01:00:00
Finish	2019/7/1 上午 08:20:00
Specification	2.0 to 8.0 °C
Average	4.3 °C
Highest	5.2 °C
Lowest	3.6 °C

Yu chen Su  
01 Jul 2019