


# TCVGH Temperature Log



2023/8/31 下午 12:00:00    2023/9/5 下午 12:00:00    2023/9/10 下午 12:00:00    2023/9/15 下午 12:00:00    2023/9/20 下午 12:00:00    2023/9/25 下午 12:00:00    2023/9/30 下午 12:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C008
Serial Number	1931-0083
Product Code	EI-IN-D-32-L
Start	2023/8/31 下午 12:00:00
Finish	2023/10/2 下午 12:00:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	4.9 °C
Highest	5.7 °C
Lowest	4.1 °C

Yuanen  
02 OCT 2023