


TCVGH Temperature Log



| | |
|---------------|-----------------------|
| Sensor Name | Internal |
| Legend | |
| Description | PHCL-C006 |
| Serial Number | 1803-0049 |
| Product Code | EI-IN-D-32-L |
| Start | 2019/8/30 下午 04:50:00 |
| Finish | 2019/10/1 上午 09:00:00 |
| Sensor Range | -40.5 to 70.4 °C |
| Specification | 2.0 to 8.0 °C |
| Average | 5.5 °C |
| Highest | 6.1 °C |
| Lowest | 5.1 °C |

Na-2 M
01 OCT 2019