


# TCVGH Temperature Log



2024/2/29	2024/3/5	2024/3/10	2024/3/15	2024/3/20	2024/3/25	2024/3/30
下午 02:00:00	下午 02:00:00	下午 02:00:00	下午 02:00:00	下午 02:00:00	下午 02:00:00	下午 02:00:00
Sensor Name	Internal					
Legend						
Description	PHCL-C006					
Serial Number	1803-0049					
Product Code	EI-IN-D-32-L					
Start	2024/2/29 下午 02:00:00					
Finish	2024/4/1 上午 08:20:00					
Sensor Range	-40.5 to 70.4 °C					
Specification	2.0 to 8.0 °C					
Average	4.9 °C					
Highest	5.9 °C					
Lowest	4.3 °C					

Yuanwen  
01 Apr 2024