


TCVGH Temperature log



| | |
|---------------|------------------------|
| Sensor Name | Internal |
| Legend | |
| Description | PHCL-C012 |
| Serial Number | 1803-0047 |
| Product Code | EI-IN-D-32-L |
| Start | 2023/10/31 下午 03:00:00 |
| Finish | 2023/12/1 上午 08:50:00 |
| Sensor Range | -40.5 to 70.4 °C |
| Specification | 2.0 to 8.0 °C |
| Average | 4.9 °C |
| Highest | 5.8 °C |
| Lowest | 4.0 °C |

Uuwennang
31 Dec 2023