


# TCVGH Temperature Log

Internal (°C)



2018/8/31 上午 10:20:00	2018/9/5 上午 10:20:00	2018/9/10 上午 10:20:00	2018/9/15 上午 10:20:00	2018/9/20 上午 10:20:00	2018/9/25 上午 10:20:00	2018/9/30 上午 10:20:00
Sensor Name Internal						
Legend						
Description PHCL-C006						
Serial Number 1803-0049						
Product Code EI-IN-D-32-L						
Start 2018/8/31 上午 10:20:00						
Finish 2018/10/1 上午 08:30:00						
Specification 2.0 to 8.0 °C						
Average 6.3 °C						
Highest 7.1 °C						
Lowest 5.4 °C						

Yu Chen Su  
01 Oct 2018