


TCVGH Temperature Log



2018/10/31 下午 03:10:00 2018/11/5 下午 03:10:00 2018/11/10 下午 03:10:00 2018/11/15 下午 03:10:00 2018/11/20 下午 03:10:00 2018/11/25 下午 03:10:00 2018/11/30 下午 03:10:00

Sensor Name	Internal
Legend	
Description	PHCL-C006
Serial Number	1803-0049
Product Code	EI-IN-D-32-L
Start	2018/10/31 下午 03:10:00
Finish	2018/12/3 上午 08:20:00
Specification	2.0 to 8.0 °C
Average	5.3 °C
Highest	6.1 °C
Lowest	4.5 °C

Na-zny 03 Dec 2018