


TCVGH Temperature Log



Sensor Name	Internal
Legend	
Description	PHCL-C005
Serial Number	1715-0115
Product Code	EI-IN-D-32-L
Start	2019/3/29 下午 12:00:00
Finish	2019/5/1 上午 08:30:00
Specification	2.0 to 8.0 °C
Average	4.9 °C
Highest	5.8 °C
Lowest	4.4 °C

Nā-zhū 01 May 2019