


TCVGH Temperature Log

Internal (°C)



2019/5/31 下午 12:00:00 2019/6/5 下午 12:00:00 2019/6/10 下午 12:00:00 2019/6/15 下午 12:00:00 2019/6/20 下午 12:00:00 2019/6/25 下午 12:00:00 2019/6/30 下午 12:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C005
Serial Number	1715-0115
Product Code	EI-IN-D-32-L
Start	2019/5/31 下午 12:00:00
Finish	2019/7/1 上午 08:30:00
Specification	2.0 to 8.0 °C
Average	4.9 °C
Highest	5.7 °C
Lowest	4.4 °C

Yu chen Su
01 Jul 2019