


TCVGH Temperature Log



Sensor Name	Internal
Legend	
Description	PHCL-C008
Serial Number	1931-0074
Product Code	EI-IN-D-32-L
Start	2021/5/31 上午 09:30:00
Finish	2021/7/1 上午 08:30:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	6.1 °C
Highest	8.5 °C
Lowest	5.5 °C

Yuchen Su 01 Jul 2021

The temperature excursion at site due to freezer filter cleaning
 The temperature excursion occurred (Max: 8.5) at 12:10 PM on 04 June 2021
 and returned to normal range (7.2) at 12:20 PM