


TCVGH Temperature Log



2021/1/29 下午 12:00:00 2021/2/3 下午 12:00:00 2021/2/8 下午 12:00:00 2021/2/13 下午 12:00:00 2021/2/18 下午 12:00:00 2021/2/23 下午 12:00:00 2021/2/28 下午 12:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C006
Serial Number	1715-0134
Product Code	EI-IN-D-32-L
Start	2021/1/29 下午 12:00:00
Finish	2021/3/2 上午 08:40:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	5.6 °C
Highest	6.3 °C
Lowest	4.9 °C

Yu Chen Sy
02/ Mar / 2021