

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C005 (Model: RAO4000)

Thermograph: ESCORT,1051-0078

Date (dd-mmm-yy)	Current (AM)			Current (PM)			Max. & Min.			Comments
	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
01-Feb-17	08:30	4.4	NTH	13:00	4.7	NTH	5.1	4.3	NTH	
02-Feb-17	08:10	4.8	NTH	17:17	4.6	NTH	5.7	4.3	NTH	
03-Feb-17	08:20	4.9	NTH	17:00	4.4	STW	5.2	4.3	NTH	
06-Feb-17	08:58	4.3	NTH	17:15	4.6	STW	5.1	4.3	NTH	
07-Feb-17	08:20	4.9	NTH	17:00	4.8	STW	5.3	4.5	NTH	
08-Feb-17	08:20	4.9	NTH	17:10	4.8	STW	5.6	4.6	NTH	
09-Feb-17	08:15	5.1	NTH	17:05	5.1	STW	5.5	4.7	NTH	
10-Feb-17	08:30	5.0	NTH	17:05	4.8	STW	5.4	4.7	NTH	
13-Feb-17	08:13	5.2	NTH	17:00	5.0	STW	5.4	4.7	NTH	
14-Feb-17	08:05	5.3	NTH	17:25	5.2	NTH	5.5	4.7	NTH	
15-Feb-17	08:25	4.8	NTH	18:00	5.1	NTH	5.5	4.7	NTH	
16-Feb-17	08:50	5.3	NTH	17:30	5.0	NTH	5.4	4.9	NTH	
17-Feb-17	08:00	5.4	NTH	17:05	5.2	NTH	5.5	4.9	NTH	
20-Feb-17	08:53	5.0	NTH	18:05	5.0	NTH	5.4	4.8	NTH	
21-Feb-17	08:06	5.0	NTH	17:05	5.1	STW	5.6	4.9	NTH	
22-Feb-17	08:18	5.1	NTH	17:24	5.1	NTH	5.4	4.8	NTH	
23-Feb-17	08:50	5.0	NTH	17:00	5.1	STW	5.4	4.9	NTH	
24-Feb-17	08:50	4.9	NTH	17:00	5.1	STW	5.7	4.8	NTH	
27-Feb-17	08:23	5.0	NTH	17:12	5.1	NTH	5.3	4.7	NTH	

Nanika Hy 07 Mar 2017