

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C005 (Model: RAO4000)

Thermograph: ESCORT,1051-0078

Date (dd-mmm-yy)	Current (AM)			Current (PM)			Max. & Min.			Comments
	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
01-May-17	09:03	5.2	NTN	17:00	5.3	STW	6.3	4.9	NTN	
02-May-17	09:19	5.4	NTN	18:06	5.3	NTN	5.5	4.9	NTN	
03-May-17	8:40	5.2	NTN	17:10	5.1	STW	5.4	4.9	NTN	
04-May-17	09:15	5.0	NTN	16:55	5.3	STW	5.6	4.9	NTN	
05-May-17	10:00	5.3	NTN	17:35	5.2	NTN	5.6	4.8	NTN	
08-May-17	08:40	5.1	NTN	17:09	4.9	STW	6.3	4.8	NTN	
09-May-17	09:10	4.9	NTN	17:00	5.1	STW	5.4	4.8	NTN	
10-May-17	09:43	5.2	NTN	17:25	5.0	STW	5.4	4.8	NTN	
11-May-17	09:29	4.8	NTN	17:00	5.5	STW	5.6	4.8	NTN	
12-May-17	08:20	5.3	NTN	17:00	5.6	STW	5.8	5.2	NTN	
15-May-17	09:05	5.3	NTN	18:00	5.3	NTN	5.7	5.3	NTN	
16-May-17	09:35	5.5	NTN	17:40	5.3	NTN	5.7	5.3	NTN	
17-May-17	09:38	5.6	NTN	17:50	5.5	NTN	5.8	5.3	NTN	
18-May-17	09:05	5.5	NTN	17:20	5.4	STW	5.8	5.3	NTN	
19-May-17	08:30	5.6	NTN	17:20	5.8	NTN	6.0	5.3	NTN	
22-May-17	08:40	5.7	NTN	17:40	5.8	NTN	6.1	5.5	NTN	
23-May-17	09:15	5.7	NTN	17:40	5.5	NTN	5.8	5.5	NTN	
24-May-17	09:05	5.6	NTN	17:20	5.5	NTN	5.9	5.4	NTN	
25-May-17	09:20	5.5	NTN	18:10	5.4	NTN	5.8	5.3	NTN	
26-May-17	08:24	5.6	NTN	17:00	5.4	STW	5.8	5.4	NTN	
29-May-17	08:31	5.7	NTN	18:05	5.6	NTN	5.8	5.4	NTN	
31-May-17	09:05	5.7	NTN	17:00	5.3	STW	5.8	5.1	NTN	

Na-NTN 01 Jun 2017