

Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2~8°C)

PHCL-C005 (Model: RAO4000)

Thermograph: ESCORT,1051-0078 (01 Jun 2017 - 05 Jun 2017) ; 1045-0031 (05 Jun 2017 - 03 Jul 2017)

Date (dd-mmm-yy)	Current (AM)			Current (PM)			Max. & Min.			Comments
	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
01-Jun-17	08:46	5.3	~YH	17:38	5.2	~YH	6.3	4.7	~YH	
02-Jun-17	09:25	4.9	~YH	17:12	4.9	STHW	5.3	4.9	~YH	
05-Jun-17	08:36	5.2	~YH	17:00	5.7	STHW	6.7	4.9	~YH	1045-0031
06-Jun-17	09:05	5.4	~YH	17:00	4.9	STHW	5.7	4.7	~YH	
07-Jun-17	08:15	5.2	~YH	17:25	5.3	STHW	5.5	4.7	~YH	
08-Jun-17	08:44	4.7	~YH	17:40	5.3	~YH	5.6	4.7	~YH	
09-Jun-17	08:15	5.3	~YH	17:05	5.3	~YH	5.7	5.1	~YH	
12-Jun-17	08:20	5.4	~YH	17:00	5.6	STHW	5.7	5.1	~YH	
13-Jun-17	08:17	5.2	~YH	17:35	5.2	STHW	5.9	4.9	~YH	
14-Jun-17	09:35	5.3	~YH	17:10	5.3	STHW	5.7	5.1	~YH	
15-Jun-17	09:15	5.2	~YH	17:00	5.1	STHW	5.7	5.1	~YH	
16-Jun-17	08:38	5.6	~YH	17:00	5.1	STHW	5.7	5.0	~YH	
19-Jun-17	08:15	5.5	~YH	17:05	5.1	STHW	5.8	4.9	~YH	
20-Jun-17	08:50	5.1	~YH	17:25	5.1	~YH	5.6	4.9	~YH	
21-Jun-17	08:35	5.3	~YH	17:00	4.8	STHW	5.4	4.7	~YH	
22-Jun-17	09:16	4.8	~YH	17:00	4.9	STHW	5.3	4.7	~YH	
23-Jun-17	08:55	4.9	~YH	17:20	5.3	STHW	5.6	4.8	~YH	
26-Jun-17	08:15	5.0	~YH	17:25	4.9	~YH	5.3	4.7	~YH	
27-Jun-17	08:15	5.0	~YH	17:30	5.2	~YH	5.3	4.7	~YH	
28-Jun-17	08:25	5.3	~YH	17:00	5.1	STHW	5.3	4.8	~YH	
29-Jun-17	09:10	4.8	~YH	17:00	5.1	STHW	5.5	4.7	~YH	
30-Jun-17	09:45	4.8	~YH	17:35	4.9	~YH	5.4	4.7	~YH	

Nairany 03 Jul 2017