

# Fridge Temperature Log of VGHTC Clinical Trial Pharmacy

(Normal Range: 2 ~ 8°C)

PHCL-C005 ( Model: RAO4000 )

Thermograph: ESCORT,1051-0078

Date (dd-mmm-yy)	Current (AM)			Current (PM)			Max. & Min.			Comments
	Time (hh:mm)	Temp (°C)	Signature (Initials)	Time (hh:mm)	Temp (°C)	Signature (Initials)	Max.Temp (°C)	Min.Temp (°C)	Signature (Initials)	
03-Jul-17	09:12	4.7	~YH	17:05	5.1	STW	5.8	4.6	~YH	
04-Jul-17	09:07	4.9	~YH	17:30	4.8	~YH	5.3	4.7	~YH	
05-Jul-17	08:20	4.7	~YH	17:37	5.1	STW	5.2	4.6	~YH	
06-Jul-17	09:15	4.8	~YH	18:00	4.8	~YH	5.3	4.7	~YH	
07-Jul-17	08:25	4.7	~YH	17:25	4.9	STW	5.2	4.6	~YH	
10-Jul-17	08:24	4.9	~YH	17:05	4.8	STW	5.4	4.7	~YH	
11-Jul-17	09:05	4.7	~YH	18:15	5.0	~YH	5.4	4.7	~YH	
12-Jul-17	08:25	4.7	~YH	17:10	5.2	STW	5.3	4.7	~YH	
13-Jul-17	09:05	5.1	~YH	17:00	4.9	STW	5.4	4.7	~YH	
14-Jul-17	09:20	5.2	~YH	17:15	5.2	STW	5.4	4.8	~YH	
17-Jul-17	08:18	5.3	~YH	17:00	5.1	STW	5.3	4.9	~YH	
18-Jul-17	09:15	5.2	~YH	17:05	5.3	STW	5.5	4.9	~YH	
19-Jul-17	09:07	5.1	~YH	17:05	4.8	STW	5.4	4.7	~YH	
20-Jul-17	09:18	4.8	~YH	17:05	5.1	STW	5.4	4.7	~YH	
21-Jul-17	09:15	5.3	~YH	17:30	5.4	~YH	5.4	4.9	~YH	
24-Jul-17	08:05	5.3	STW	17:00	5.0	STW	5.6	4.9	~YH	
25-Jul-17	09:40	5.3	~YH	17:05	5.3	STW	5.8	4.8	~YH	
26-Jul-17	09:15	5.1	~YH	17:25	5.6	~YH	5.6	4.8	~YH	
27-Jul-17	08:30	4.9	~YH	16:55	5.3	STW	5.7	4.9	~YH	
28-Jul-17	09:20	5.2	~YH	17:00	5.2	STW	5.8	5.1	~YH	
31-Jul-17	08:25	5.2	~YH	17:05	5.2	STW	5.6	5.1	~YH	

Na-6 14/02 Aug 2017