


TCVGH Temperature Log



Sensor Name	Internal
Legend	
Description	PHCL-C003
Serial Number	1715-0143
Product Code	EI-IN-D-32-L
Start	2021/9/30 下午 02:00:00
Finish	2021/11/1 上午 08:20:00
Sensor Range	-40.5 to 70.4 °C
Specification	2.0 to 8.0 °C
Average	6.0 °C
Highest	8.3 °C
Lowest	4.7 °C

The Temperature excursion occurred due to long time opening the door of refrigerator when placing plasma other IPs on 22 Oct 2021 at 17:20 max= 8.3°C ,
 45, 011111
 Another event was occurred during 17:50-18:00 when taking out the temperature device to download the temperature recorder .
 11111111 01111111