


TCVGH Temperature Log



2018/4/30 下午 07:20:00 2018/5/5 下午 07:20:00 2018/5/10 下午 07:20:00 2018/5/15 下午 07:20:00 2018/5/20 下午 07:20:00 2018/5/25 下午 07:20:00 2018/5/30 下午 07:20:00

Sensor Name	Internal
Legend	
Description	PHCL-C003
Serial Number	1803-0048
Product Code	EI-IN-D-32-L
Start	2018/4/30 下午 07:20:00
Finish	2018/6/1 上午 08:10:00
Specification	2.0 to 8.0 °C
Average	3.9 °C
Highest	5.3 °C
Lowest	2.4 °C

Noted by 01 Jun 2018