


TCVGH Temperature log



2018/5/31	2018/6/5	2018/6/10	2018/6/15	2018/6/20	2018/6/25	2018/6/30
上午 09:20:00	上午 09:20:00	上午 09:20:00	上午 09:20:00	上午 09:20:00	上午 09:20:00	上午 09:20:00

Sensor Name Internal
 Legend
 Description PHCL-C003
 Serial Number 1715-0143
 Product Code EI-IN-D-32-L
 Start 2018/5/31 上午 09:20:00
 Finish 2018/7/2 上午 08:30:00
 Specification 2.0 to 8.0 °C
 Average 4.0 °C
 Highest 5.2 °C
 Lowest 2.1 °C

Yu wen wang 02 Jul 2018