


# TCVGH Temperature Log



2018/8/31 上午 10:20:00	2018/9/5 上午 10:20:00	2018/9/10 上午 10:20:00	2018/9/15 上午 10:20:00	2018/9/20 上午 10:20:00	2018/9/25 上午 10:20:00	2018/9/30 上午 10:20:00
Sensor Name Internal						
Legend						
Description	PHCL-C003					
Serial Number	1803-0048					
Product Code	EI-IN-D-32-L					
Start	2018/8/31 上午 10:20:00					
Finish	2018/10/1 上午 08:20:00					
Specification	2.0 to 8.0 °C					
Average	3.9 °C					
Highest	5.1 °C					
Lowest	3.1 °C					

Yu Chen Su  
01 Oct 2018