


TCVGH Temperature Log



2018/9/28 下午 01:00:00 2018/10/5 下午 01:00:00 2018/10/12 下午 01:00:00 2018/10/19 下午 01:00:00 2018/10/26 下午 01:00:00

Sensor Name	Internal
Legend	
Description	PHCL-C003
Serial Number	1715-0143
Product Code	EI-IN-D-32-L
Start	2018/9/28 下午 01:00:00
Finish	2018/11/1 上午 09:10:00
Specification	2.0 to 8.0 °C
Average	4.1 °C
Highest	5.4 °C
Lowest	3.2 °C

Na-zny 01 Nov 2018